


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AV4-MPT-F*

S.N.: *6148*

Date: *JAN 30 1992*



*40 dB ATTEN*

*10 VOLTS/DIV*

*PRF = 700 kHz*

a) Output signal amplitude:

*≥ 30 VOLTS*

b) Pulse width:

*≤ 1 nS*

c) Rise time:

*NA*

d) Fall time:

*NA*

e) PRF:

*0 TO 1.0 MHz*

f) Jitter, stability:

*OK*

g) Prime power:

*+15 VOLTS,*

*100 mA (MAX)*