


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVA-MPT-F*

S.N.: *6149*

Date: *JAN 30 1992*



40 db ATTEN

10 VOLTS/DIV

PRF = 700 kHz

- a) Output signal amplitude:
> 30 VOLTS
- b) Pulse width:
≤ 1 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 10 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*+15 VOLTS
100 mA (max)*

[Signature]