


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVL-3AH-PS-P-T-M*

S.N.: *6161*

Date: *MAR 10 1992*




Ⓐ 70 db ATTN  
∴ 320 VOLTS/DIV  
MIN TR



Ⓑ AS Ⓐ BUT  
MIN TR

- a) Output signal amplitude:  
*0 TO +450 VOLTS*
- b) Pulse width: *TO 50 ns*
- c) Rise time:  
*≤ TO 100 NS*
- d) Fall time:  
*≤ 2 NS TO 10 NS*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*120 / 240 V*  
*50 - 80 W*



Ⓒ AS Ⓐ BUT TRI AND TFI  
CONTROLS NOT SET  
PROPERLY (TO GIVE)  
MINIMUM RISE, FALL

*BA*