


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVM-3-P*

S.N.: *6169*

Date: *FEB 6 1992*



40 dB ATTEN

10 VOLTS/DIV

PRF = 10 MHz

- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
2 TO 15 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 600 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+20 VOLTS DC
500 mA (MAX)