


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-53-PS-R4-PEC2

S.N.: 6171

Date: FEB 28 1992



a) Output signal amplitude:
120 VOLTS

b) Pulse width:
TDF: 125 NS

c) Rise time:
SALI: 12 NS
TDF: 1 NS


d) Fall time:
SALI: 20 NS
TDF: 1 NS

e) PRF: SFR: 20 NS
TDF: 0 TO 20 KHz


f) Jitter, stability:
SALI: 0 TO 1 KHz

g) Prime power:
OK
120/240 VOLTS
50-60 Hz


ⓐ TOFA 100V/DIV (60db)



ⓑ TOFB 100V/DIV (60db)



ⓒ SALI A




ⓓ SALI A

PULSE GENERATOR
PERFORMANCE CHECK

Model:

S.N.: 6171 CONT.

Date:



a) Output signal amplitude:

b) Pulse width:

c) Rise time:


d) Fall time:

e) PRF:

f) Jitter, stability:

g) Prime power:

⑤ 5A21 B



20V

⑤ 5A21 B