


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PR-A1-PW-C-N-FL1*

S.N.: *6176*

Date: *JUN 28 1992*



R_L = 50 Ω
PRF = 50 kHz

- a) Output signal amplitude:
0 TO -200 VOLTS
- b) Pulse width: *(TO 50 μ s)*
50 TO 100 NS
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF: *50 Hz TO 50 kHz*
- f) Jitter, stability:
OK
- g) Prime power:
120/200 V
50-60 Hz.