


PULSE GENERATOR

PERFORMANCE CHECK


Model: AVG-2-C-7-SNLI-TPI-R4

S.N.: 6180


Date: FEB 21 1992



RANGE 1 60 db ATTEN
100 VOLTS/DIV
PRF = 10 KHz



RANGE 2 60 db ATTEN
100 VOLTS/DIV
PRF = 10 KHz



RANGE 3 60 db ATTEN
100 VOLTS/DIV
PRF = 10 KHz

a) Output signal amplitude:

0 TO +100 VOLTS

b) Pulse width:

2, 4, 10 + 20 NS

c) Rise time:

NA

d) Fall time:

NA

e) t_p : < 40 NS

e) PRF:


10 Hz TO 10 KHz

f) Jitter, stability:

OK

g) Prime power:

120 / 240 V
50 - 60 Hz



RANGE 4 60 db ATTEN
100 VOLTS/DIV
PRF = 10 KHz