


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-14V1-PN-55C1*

S.N.: *6186*


Date: *FEB 6 1992*



- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
P: 0.6 TO 2.0 NS
N: 0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA

Pass 60 db - 100 VOLTS/DIV
30 KHz

- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability:
OK



- g) Prime power:
+15 VOLTS DC
150 mA (MAX)

[Signature]

Now 60db - 100 VOLTS/DIV
30 KHz