


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AMP-MV-1-CT*

S.N.: *6192*

Date: *FEB 17 1992*



*40 dB ATTEN*

*10 VOLTS/DIV*

*PRF = 100 MHz*

- a) Output signal amplitude:  
*0 TO +10 VOLTS*
- b) Pulse width:  
*0.2 TO 4.0 NS*
- c) Rise time:  
*≤ 100 PS*
- d) Fall time:  
*≤ 135 PS*
- e) PRF:  
*0 TO 1.0 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*120/240 V*  
*50-60 Hz*