


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVH-MVI-PN-55C1*

S.N.: *6196*

Date: *MAR 16 1992*



*P_{out} 60 db
- 100 VOLTS/DIV
60 KHz*



*N_{out} 60 db
100 VOLTS/DIV
60 KHz.*

- a) Output signal amplitude:
± 100 VOLTS
- b) Pulse width:
PN: 0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:
*+15 VOLTS DC
150 mA (MAX)*

[Signature]