


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-5-1-P

S.N.: 6200

Date: MAR 5 1992



40 dB ATTEN

10 VOLTS/DIV

PRF ≈ 300 kHz

- a) Output signal amplitude: 0 TO +10 VOLTS
- b) Pulse width: TO 50 ns
≤ 130 ps
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 10 MHz
- f) Jitter, stability: OK
- g) Prime power: +15 VOLTS, 150 mA (MAX)