


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVH-5-P-GSSI*

S.N.: *6201*

Date: *MAR 5 1992*



*40dB ATTEN*  
*10 VOLTS/DIV*  
*PRF ≈ 300 KHz*

- a) Output signal amplitude:  
*0 TO +15 VOLTS*
- b) Pulse width:  
*≤ 500 PS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF: *0 TO 1.0 MHz.*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 VOLTS*  
*150 mA (MAX)*

