


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVR-51-C-PWT-AT-PN*

S.N.: *6209*


Date: *MMAR 25 1992*



*P_{out} R_L = 50 Ω
PRF = 1 kHz*




P_{out} 100 VOLTS/DIV (RISE TIME)



N_{out} 100 VOLTS/DIV (RISE TIME)

- a) Output signal amplitude: *0 TO ±200 VOLTS TO 50 Ω*
- b) Pulse width: *0.1 TO 5.0 μS*
- c) Rise time: *≤ 2 NS*
- d) Fall time: *≤ 10 NS*
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability: *OK*
- g) Prime power: *120/240 V
50-60 Hz*



*N_{out} R_L = 50 Ω
PRF = 1 kHz*