


PULSE GENERATOR

PERFORMANCE CHECK


Model: *PH-141-PN-SSC1*

S.N.: *6221*

Date: *MAR 30 1992*



*Fout 60db*  
*∴ 100 VOLTS/DIV*  
*30 KHZ*



*Fout 60db*  
*100 VOLTS/DIV*  
*30 KHZ*

- a) Output signal amplitude:  
*± 100 VOLTS*
- b) Pulse width:  
*PWN: 0.6 TO 2.0 NS*
- c) Rise time:
- d) Fall time: *NA*
- e) PRF: *NA*
- f) Jitter, stability:  
*0 TO 100 KHZ.*
- g) Prime power:  
*+15 VOLTS DC*  
*150 mA (MAX)*