


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PH-141-PN-55C1*

S.N.: *6227*

Date: *MAR 30 1992*



a) Output signal amplitude:

*± 100 VOLTS*

b) Pulse width:

*PWN: 0.6 TO 2.0 NS*

c) Rise time:

*NA*

d) Fall time:

*NA*

e) PRF:

*0 TO 100 KHZ*

f) Jitter, stability:


*6K*

g) Prime power:

*+15 VOLTS DC.  
150 mA (MAX)*

*[Signature]*

*P<sub>OUT</sub> 60 db  
100 VOLTS/PIU  
30 KHZ*



*N<sub>OUT</sub> 60 db  
100 VOLTS/PIU  
30 KHZ*