


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AP-M-1A3-C-P*

S.N.: *6231*

Date: *MAR 31 1992*



50 dB ATTEN
32 VOLTS/DIV
PRF = 100 KHz

- a) Output signal amplitude:
0 TO +40 VOLTS
- b) Pulse width: (*TO 50ns*)
0.4 TO 2.0 NS
- c) Rise time:
≤ 150 PS
- d) Fall time:
≤ 250 PS
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS,
50-60 Hz.

[Signature]