


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-6A-C-P-ET-EN-LIB*

S.N.: *6237*

Date: *APRIL 16 1992*



- a) Output signal amplitude:
0 TO + 1.0 AMP
- b) Pulse width:
50 NS TO 5.0 U)
- c) Rise time:
≤ 10 NS
- d) Fall time:
≤ 10 NS
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*MAIN FRAME OUT
TO $R_L = 50 \Omega$.*

*120/240 V
50 = 60 Hz*

[Signature]