


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MH-171-P*

S.N.: *6240*

Date: *APRIL 30 1992*



60 db ATTEN
100 VOLTS/DIV
PRF = 30 KHz

- a) Output signal amplitude:
0 TO 7100 VOLTS
- b) Pulse width: *(to 50%)*
≤ 1.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF: *0 TO 100 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS DC
250 mA (MAX)