


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVR-EI-C-P*  
S.N.: *6250*  
Date: *APR 5 1992*



*10 TOP*  
*VOLTS/DIV*  
*20 NS/DIV*  
  
*BOT*  
*10 VOLTS/DIV*  
*500 PS/DIV*  
*(RISE TIME)*

- a) Output signal amplitude: *0 TO 720 VOLTS TO*
- b) Pulse width: *50  $\mu$ s*
- c) Rise time: *20 TO 200 NS*
- d) Fall time:  *$\leq$  300 PS*
- e) PRF:  *$\leq$  300 PS*
- f) Jitter, stability: *0 TO 100 KHz.*
- g) Prime power: *OK*
- h) *120 / 240 V*
- i) *50 - 60 Hz.*