


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-2W.-CT*

S.N.: *6253*

Date: *APRIL 1 1992*



1/2 7 AMPS/DIV

PRF = 10 KHz

- a) Output signal amplitude:
0 TO +10 AMPS
- b) Pulse width:
2 TO 58 NS
- c) Rise time:
≤ 1 NS
- d) Fall time:
≤ 2 NS
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*120/240 VOLT,
50-60 Hz.*