


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ANM-1-C-PN*

S.N.: *6259*

Date: *APRIL 13 1992*



- a) Output signal amplitude:
0 TO ± 5 VOLTS
- b) Pulse width:
0.2 TO 6.0 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK

40 dB ATTEN
- 5 VOLTS/DIV
PRF ≈ 12 MHz.

g) Prime power:
120/240V
50-60Hz
[Signature]