


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PVI-MAP-P*
S.N.: *6269*
Date: *MAY 4 1992*



- a) Output signal amplitude:
> +40 VOLTS TO
- b) Pulse width: *50 n*
2 TO 100 NS
- c) Rise time:
≤ 350 ps.
- Fall time:
62 NS
- PRF:
PW MIN: 0 TO 1 MHz
PW MAX: 0 TO 20 kHz
- f) Jitter, stability: *SEE FIG 1 + 2*
- g) Prime power:
OK

+15 VOLTS DC
100 mA (MAX)