


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-AU-C-N-EA-PL1-05

S.N.: 6273

Date: JULY 13 1992



60 dB ATTEN

100 VOLTS/DIV

PRF = 2.0 KHZ

- a) Output signal amplitude:
0 TO 250 VOLTS
- b) Pulse width: 70 50ns
8 TO 100 NS
- c) Rise time:
 ≤ 5 NS
- d) Fall time:
 ≤ 5 NS
- e) PRF: 0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
120/240 V
50-60 HZ

h) DC OFFSET:

PL1-05 OFFSET

0 to ± 250 VOLTS

100 mA MAX.