


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AV6-1-P-LT1*
S.N.: *6277*
Date: *MAY 4 1992*



60 db return

100 VOLTS/DIV

PRF = 30 KHZ

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width: *TO 50ns*
≤ 4 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 KHZ
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS DC
250mA (MAX)