


PULSE GENERATOR
PERFORMANCE CHECK

Model: AV6-1-P-LT1

S.N.: 6278

Date: MAR 4 1992



- a) Output signal amplitude:
0 TO + 250 VOLTS
- b) Pulse width: TO 50ns
64 ns
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:

60 db ATTEN
100 VOLTS/DIV
PRF = 30 kHz

+15 VOLTS DC
250 mA (MAX)