


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVH-AVI-PN-55C1*

S.N.: *6280*

Date: *APRIL 28 1992.*



*P_{out} 60 db
100 VOLTS*



N_{out} 143 P.

- a) Output signal amplitude:
±100 VOLTS
- b) Pulse width:
≤ 0.6 TO 2.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 100 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*+15 VOLTS,
150 mA MAX.*

- h) SPURIOUS:
*≈ x 10⁻¹⁰ LOWER
THAN FIRST 8
UNITS IN
THIS SERIES*

A