



PULSE GENERATOR
PERFORMANCE CHECK

Model: AVI-V-P
 S.N.: 6316
 Date: MAY 29 92



ⓐ 50 dB ATTEN
 * 32 VOLTS/DIV
 PRF = 20 KHz



ⓑ AS ⓐ BUT
 500 PS / DIV
 (RISE TIME)

- a) Output signal amplitude:
0 TO +50 VOLTS
- b) Pulse width: (70 50ns)
2 TO 100 NS
- c) Rise time:
≤ 350 PS
- d) Fall time:
≤ 2 NS
- e) PRF: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:

+15 VOLTS
 250 mA
 (MAX)

[Handwritten signature]