


PULSE GENERATOR
PERFORMANCE CHECK


Model: *AVL-3A-C-P*

S.N.: *6317*

Date: *MAY 21 1992*



*ⓐ 70 db ATTEN
320 VOLTS/DIV
PRF ≈ 2.0 KHz*



*ⓑ AS ⓐ BUT 500ps/DIV
(RISE TIME CHECK)*

- a) Output signal amplitude:
0 TO +300 VOLTS
- b) Pulse width: *70 50 ns*
5 TO 100 ns
- c) Rise time:
≤ 600 ps
- d) Fall time:
≤ 3 ns
- e) PRF:
0 TO 5 KHz.
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS
50-60 MHz

LA