


PULSE GENERATOR
PERFORMANCE CHECK

Model: *ML-2-P*

S.N.: *6318*

Date: *MAY 22 1992*



70 dB ATTEN

~ 320 VOLTS/DIV

PRF = 3.0 KHz.

- a) Output signal amplitude:
0 TO +350 VOLTS
- b) Pulse width: *(TO 50ns)*
- c) Rise time: *5 TO 100 NS*
≤ 2 NS
- d) Fall time: *≤ 3 NS*
- e) PRF: *0 TO 5 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
500 mA (max)

A