


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVL-AV-1-P

S.N.: 6319

Date: MAY 22 1992



- a) Output signal amplitude:  
0 TO +100 VOLTS
- b) Pulse width:  
3 TO 100 NS
- c) Rise time:  
 $\leq 1$  NS
- d) Fall time:  
 $\leq 2$  NS
- e) PRF:  
0 TO 5 KHz.
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VOLTS  
500 mA (MAX)

60 db  
ATTEN

100 VOLTS/DIV

PRF = 5.0 KHz.

(1.0 NS/DIV)  
RISE  
TIME