


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVI-U-HV1-C-EA EW-PN*

S.N.: *6332*

Date: *JUNE 11 1992*



a) Output signal amplitude:  
*0 TO ±50 VOLTS*

b) Pulse width:  
*5 TO 100 NS*

c) Rise time:  
*≤ 250 PS*

*← 10 NS/DIV*  
d) Fall time:  
*≤ 2 NS*

e) PRF:  
*0 TO 20 KHZ*

f) Jitter, stability:

*(RISE TIME)*  
*OK*

g) Prime power:

*120 / 240 V*  
*50 - 60 Hz*

*50 dB ATTEN*

*32 VOLTS/DIV*