


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-3A-P-LLA

S.N.: 6333

Date: June 23 1992



70 dB ATTEN

$\therefore \approx 3.20 \text{ VOLTS/DIV}$

$\text{PRF} \approx 10 \text{ KHz}$

- a) Output signal amplitude:
 $\approx +300 \text{ VOLTS/DIV}$
- b) Pulse width: 70 ns
 10 ns
- c) Rise time:
 $\leq 600 \text{ ps}$
- d) Fall time:
 $\approx 4 \text{ ns}$
- e) PRF:
 $0 \text{ TO } 20 \text{ KHz}$
- f) Jitter, stability:
 OK
- g) Prime power:
 $+15 \pm 1.0 \text{ VOLT}$
 250 mA (MAX)

[Signature]