


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-3A-P-LLA

S.N.: 6334

Date: June 23 1992



70 db ATTEN

≈ 320 VOLTS/DIV

PRF ≈ 10 KHz

- a) Output signal amplitude:
 $\approx +300$ VOLTS
- b) Pulse width: ≈ 50 ns
 10 NS
- c) Rise time:
 ≤ 600 PS
- d) Fall time:
 ≈ 4 NS
- e) PRF: 0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
 $+15 \pm 1.0$ VOLTS
 250 mA (MAX)

[Signature]