


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PH-0MPT-TP1-P*

S.N.: *6337*

Date: *MAY 30 1992*



- a) Output signal amplitude:
≥ +30 VOLTS
- b) Pulse width:
≤ 1.0 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
100 mA (PULSED)
- h) t_p : *≤ 2 NS.*

40 db ATTEN
10 VOLTS/DIV
PRF = 1.0 MHz