


PULSE GENERATOR
PERFORMANCE CHECK


Model: *PK-7B-C-PN-CA-CA-D-6W*
-AT-PWT

S.N.: *6342*

Date: *JUNE 25 1992*



Pout R_L = 50Ω
PRF = 5 kHz



Nout R_L = 50Ω
PRF = 50 kHz

- a) Output signal amplitude: *0 TO ±700 VOLTS*
- b) Pulse width: *TO R_L 750Ω*
- c) Rise time: *0.1 TO 100 ns*
0.5% MARR
DUTY cycle
- d) Fall time: *±50 ns*
- e) PRF: *0 TO 10 kHz*
- f) Jitter, stability: *0.5% MARR*
DUTY cycle
- g) Prime power: *OK*
120/240 VOLTS
50-60 Hz
LM