


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVG-1-P

S.N.: 6344

Date: JUNE 12, 1992



60 db ATTEN

\therefore 100 VOLTS/DIV

PRF = 30 kHz

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width:
 ≤ 4 NS
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 50 kHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 VOLTS
250 mA