


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVG-1-P

S.N.: 6346

Date: JUNE 12, 1992



60 db ATTEN

∴ 100 VOLTS/DIV

PRF = 30 kHz

- a) Output signal amplitude:  
0 TO +250 VOLTS
- b) Pulse width:  
≤ 4 NS
- c) Rise time:  
NA
- d) Fall time:  
NA
- e) PRF:  
0 TO 50 kHz
- f) Jitter, stability:  
OK
- g) Prime power:  
+15 VOLTS  
250 mA