


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVM-1-IP*

S.N.: *6356*

Date: *SEPT 3 1992*



a) Output signal amplitude:

0 TO +5 VOLTS

b) Pulse width:

0.2 TO 6.0 NS

c) Rise time:

≤ 100 PS

d) Fall time:

≤ 135 PS

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24 VOLTS

400 mA (MAX)

40 dB ATTEN

5 VOLTS/DIV.

[Signature]