


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-S-1-C-P
S.N.: 6357
Date: JUNE 26 1992



- a) Output signal amplitude: 0 TO +10 VOLTS
- b) Pulse width: TO 50 nS
≤ 30 pS.
- c) Rise time: NA
- d) Fall time: NA
- e) PRF: 0 TO 1.0 MHz
- f) Jitter, stability: OK
- g) Prime power: 120/200 V
50-60 Hz

40 dB ATTEN
10 VOLTS/DIV.

PRF = 1.0 MHz