


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVR-A-1-PW-CR1

S.N.: 6360

Date: July 17 1992



60 dB ATTEN

100 VOLTS/DIV

PRF = 50 KHz

- a) Output signal amplitude:
0 TO +250 VOLTS
- b) Pulse width: TO 50ns
30 TO 50 NS
- c) Rise time:
≤ 5 NS
- d) Fall time:
≤ 5 NS
- e) PRF: 0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+24 VOLTS.
300 mA (MAX)

[Handwritten signature]