


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A-1-S1-C-P


S.N.: 6362

Date: JULY 16 1992



- a) Output signal amplitude: 0 TO +200 VOLTS
- b) Pulse width: 50 NS TO 500 NS
- c) Rise time: 50 NS TO 1.00 US (0.5% MAX DUTY CYCLE)
- d) Fall time: ≤ 2 NS
- e) PRF: 0 TO 100 KHZ BUT 0.5% MAX DUTY
- f) Jitter, stability: CYCLE, OK
- g) Prime power: 120/240 V, 50-60 HZ

① $R_L = 50 \Omega$
PRF = 5.0 KHZ



② RISE TIME CHECK
60 DB ATTEN
100 VOLTS/DIV

[Signature]