


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-C-P-05-M

S.N.: 6363

Date: JUL 7 1992



- a) Output signal amplitude:
0 TO +350 VOLTS
- b) Pulse width: (TO 50ns)
- c) Rise time: 5 TO 100NS
≤ 2 NS
- d) Fall time:
≤ 3 NS
- e) PRF: 0 TO 5 KHz
- f) Jitter, stability:
OK
- g) Prime power:

TO 16 ATTEN
≈ 320 VOLTS/DIV
PRF = 20 KHz

120 / 240 VOLTS
50 - 60 Hz.
[Signature]