


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVI-V-M-1-C-P-PN-ET-5W-05-A

S.N.: 6364

Date: Aug 14 1992



- a) Output signal amplitude:  
0 TO  $\pm 50$  VOLTS
- b) Pulse width:  
2 TO 100 NS
- c) Rise time:  
 $\leq 250$  PS
- d) Fall time:  
 $\leq 2$  NS
- e) PRF:  
0 TO 20 KHz
- f) Jitter, stability:  
OK
- g) Prime power:  
120/200 VOLT,  
50-60 MHz

50 dB ATTEN

32 VOLTS/DIV

B&T: 200 PS/DIV  
(RISE TIME)