


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-2S-P-05-ECL

S.N.: 6367

Date: JULY 9/92



- a) Output signal amplitude:
0 TO + 2 VOLTS
- b) Pulse width:
0.2 TO 4.0 NS
- c) Rise time:
≤ 40 PS
- d) Fall time:
≤ 135 PS
- e) PRF:
0 TO 1.0 MHZ
- f) Jitter, stability:
OK
- g) Prime power:

∴ 20 db ATTEN
2 VOLTS / DIV
PRF = 1.0 MHZ

$$T_{RT} = \sqrt{T_{RAS}^2 + T_{SCOPE}^2 + T_{ATTEN}^2}$$

$$\approx 50 \text{ PS (20-80\%)}$$

T SCOPE = 25 PS

18 GHz ATTEN + CABLES

∴ CONC THAT $T_{RAS} \leq 40 \text{ PS}$

+ 15 VOLTS
400 mA (MAX)