


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-6B-P-C-AT-USC1

S.N.: 6373

Date: July 14 1992



- a) Output signal amplitude:
0 TO +2 AMPS
- b) Pulse width:
0.1 TO 5.0 μ S
- c) Rise time:
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF: 0 TO 10 KHz
- f) Jitter, stability:
OK

2 VOLTS / DIV
200 NS / DIV

AVO-6B-M-05-USC1
MONITOR OUT TO
50 Ω .

- g) Prime power:
120 / 240 V
50 - 60 Hz

$$I_{DIPPE} = \frac{(10)(10.4) - 2}{50}$$

≈ 2.0 AMPS