


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AMP-AN-2-C-P-PN*

S.N.: *6382*

Date: *AUG 14 1992*



40 dB ATTEN

10 VOLTS/DIV

PRF = 50 KHz

- a) Output signal amplitude:
0 TO ± 10 VOLTS
- b) Pulse width:
5 TO 50 NS
- c) Rise time:
≤ 100 PS
- d) Fall time:
≤ 130 PS
- e) PRF: *0 TO 50 KHz*
- f) Jitter, stability:
OK
- g) Prime power:
120 / 280 VOLTS
50 - 60 Hz

[Signature]