


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-7B-PS-PN-MHV

S.N.: 6383


Date: JULY 22 1982



- a) Output signal amplitude:
0 TO ± 700 VOLTS
- b) Pulse width:
0.1 TO 100 μ S
- c) Rise time: (0.5% MAX DUTY CYCLE)
 ≤ 50 NS
- d) Fall time:
 ≤ 50 NS
- e) PRF: 0 TO 10 KHz
0.5% MAX DUTY
- f) Jitter, stability: CYCLE
OK
- g) Prime power:

Ⓐ $P_{out}: R_c = 50\Omega$
PRF = 5 KHz

120/240 VOLT,
50-60 Hz.



Ⓑ $N_{out}: R_c = 50\Omega$
PRF = 50 MHz

[Signature]