


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MR-A-1-PW-CN*

S.N.: *6384*

Date: *Aug 14 1992*



$R_L = 50 \Omega$
PRF = 5.0 kHz

- a) Output signal amplitude:
0 TO -200 V (4.5)
- b) Pulse width:
0.05 TO 1.0 μ S
- c) Rise time: *(0.5% MAX DUTY CYCLE)*
 ≤ 10 NS
- d) Fall time:
 ≤ 10 NS
- e) PRF:
0 TO 100 kHz
(0.5 MAX DUTY CYCLE)
- f) Jitter, stability:
OK
- g) Prime power:
120/240 VOLTS,
50-60 Hz.

A