


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVP-25-P-05-ECL

S.N.: 6385

Date: AUG 4, 1992



20 db ATTEN
∴ 2 VOLTS/DIV
PRF = 1.0 MHz

$$T_{RT} = \sqrt{T_{R2S}^2 + T_{SCOPE}^2 + T_{ATTEN}^2}$$

$$\approx 50 \text{ PS (20-80\%)}$$

$$T_{SCOPE} = 25 \text{ PS}$$

18 GHz ATTEN + CABLES

∴ CONC THAT $T_{R2S} \leq 40 \text{ PS}$

a) Output signal amplitude:
0 TO + 2 VOLTS

b) Pulse width:
1 NS/DIV 0.2 TO 4.0 NS

c) Rise time:
50 PS/DIV $\leq 40 \text{ PS}$

d) Fall time:
 $\leq 135 \text{ PS}$

e) PRF:
0 TO 1.0 MHz

f) Jitter, stability:
OK

g) Prime power:
+ 15 VOLTS
400 mA (MAX)