


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVI-V-HV2-P-CEA*

S.N.: *6400*

Date: *JULY 17 1992*



*60 db ATTEN*

*100 VOLTS/DIV*

*PRF = 20 kHz*

- a) Output signal amplitude:  
*0 TO +100 VOLTS*
- b) Pulse width: *(TO 50ns)*
- c) Rise time:  
*0 TO 20 ns*  
*≤ 2 ns*
- d) Fall time:  
*≤ 2 ns*
- e) PRF:  
*0 TO 20 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*+15 VOLTS*

*250 mA (max)*

*[Signature]*